













TO STAY CONNECTED

to the myMerlin™ for Confirm Rx™ ICM mobile app, keep your smartphone on these settings:

 = iPhone⁺ settings

 = Android⁺ settings

 Bluetooth®, cellular data/ Wi-Fi ⁺	ON
 Bluetooth®, location services & cellular data/Wi-Fi ⁺	ON
 Background app refresh	ON
 Background data usage	ON
 Automatically update apps	ON
 Automatically update apps	ON
 Font size STANDARD	ON
 Font size NORMAL	ON
 Battery Low Power Mode	OFF
 Battery Saver	OFF
 Offload unused apps	OFF
 Battery optimization for “Confirm Rx™” app	OFF

For more information about
Confirm Rx™ ICM, visit

ConfirmYourRhythm.com

**Please have the following information
available when you call:**

Confirm Rx™ ICM

serial number _____

Name of clinic

that monitors you _____

Smartphone

make and model _____

KEEPING YOUR MYMERLIN™ FOR CONFIRM Rx™ ICM MOBILE APP CONNECTED

QUICK TIPS



Abbott

One St. Jude Medical Dr., St. Paul, MN 55117 USA, Tel: 1 651 756 2000
Abbott.com

RX ONLY

Brief Summary: Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use.

Indications: The Confirm Rx™ ICM is indicated for the monitoring and diagnostic evaluation of patients who experience unexplained symptoms such as: dizziness, palpitations, chest pain, syncope, and shortness of breath, as well as patients who are at risk for cardiac arrhythmias. It is also indicated for patients who have been previously diagnosed with atrial fibrillation or who are susceptible to developing atrial fibrillation. The Confirm Rx™ ICM has not been specifically tested for pediatric use.

Contraindications: There are no known contraindications for the insertion of the Confirm Rx™ ICM. However, the patient's particular medical condition may dictate whether or not a subcutaneous, chronically inserted device can be tolerated.

Adverse Events: Possible adverse events (in alphabetical order) associated with the device, include the following: Allergic reaction, Bleeding, Chronic nerve damage, Erosion, Excessive fibrotic tissue growth, Extrusion, Formation of hematomas or cysts, Infection, Keloid formation and Migration. Refer to the User's Manual for detailed indications, contraindications, warnings, precautions and potential adverse events.

Additional information: Clinicians must log onto Merlin.net™ Patient Care Network to view transmissions from patients' Confirm Rx™ ICM. On Merlin.net™ PCN they can configure transmission schedules and enable or disable features on a patient's myMerlin™ for Confirm Rx™ ICM mobile app. Review of transmissions is dependent on the clinician and may not happen immediately following delivery of such transmissions.

Limitations: Patients may use their own Apple[‡] or Android[‡] mobile device to transmit information from their Confirm Rx™ ICM using the myMerlin™ for Confirm Rx™ mobile app. To do so the device must be powered on, app must be installed, Bluetooth[®] wireless technology enabled and data coverage (cellular or WiFi) available. The myMerlin™ for Confirm Rx™ mobile app provides periodic patient monitoring based on clinician configured settings. Data is resent if the transmission was not sent successfully. However, there are many internal and external factors that can hinder, delay, or prevent acquisition and delivery of ICM and patient information as intended by the clinician. These factors include: patient environment, data services, mobile device operating system and settings, ICM memory capacity, clinic environment, schedule/configuration changes, or data processing.

An Abbott mobile transmitter is available for patients without their own compatible mobile device.

™ Indicates a trademark of the Abbott group of companies.

‡ Indicates a third-party trademark, which is property of its respective owner.

Bluetooth and the Bluetooth logo are registered trademarks of Bluetooth SIG, Inc.

© 2019 Abbott. All Rights Reserved.

SJM-CFM-0519-0134 | Item approved for U.S. and OUS use.

myMerlin™ for Confirm Rx™ ICM mobile app

If you're experiencing issues with the mobile app, contact Abbott Remote Care Technical Support:

UNITED STATES Hours of support

M-F 8a-8p (ET)

+1.877.696.3754

myMerlin@abbott.com

INTERNATIONAL Need help?

Contact your clinic directly with any questions.

You can also email
Remote.Monitoring
@abbott.com